

## TEST REPORT N° 5\_24/03/21

Issue date 24/03/2021

Esteemed Company  
 Danish Technological Institute  
 Gregersensvej, 1  
 00000 Taastrup DK-2630 (-)

**Sample type** Materials  
**Sample received on** 18/03/2021  
**Sample description** CUBO medical.clean 16.03.21<sup>1</sup>  
**Sampling site** Sampling performed at the Customer's premises<sup>1</sup>  
**Sampler** Client<sup>1</sup>  
**Sampling method** Internal to the Client<sup>1\*\*</sup>  
**Sample pack** Sample packed in plastic bag  
**Sample Condition / Seals** Sample delivered in a manner and quantity suitable for carrying out the required analytical investigations.  
**Transport by** Courier service  
**Temperature** ---

**Sample Protocol** 06\_180321 del 18/03/21

**Description** CUBO medical.clean 16.03.21

Testing Start date - End date	Result	U.M	Method	EN 14683:2019 Table1		
				I	II	IIR
Bacterial Filtration Efficiency 19/03/2021 - 21/03/2021	99,8	%	EN 14683:2019/AC 2019 App B	≥95	≥98	≥98
Negative Control	0					
1) Positive Control	2313	UFC				
2) Positive Control	2223	UFC				
1) BFE	99,8	%		≥95	≥98	≥98
2) BFE	99,9	%		≥95	≥98	≥98
3) BFE	99,9	%		≥95	≥98	≥98
4) BFE	99,9	%		≥95	≥98	≥98
5) BFE	99,9	%		≥95	≥98	≥98

### Additional information

The analytical determinations were performed on 5 specimens, cut from complete masks / original fabric that makes up the mask.  
 Each sample is 100mm × 100mm in size and includes all mask layers in the order they are inserted into the full mask.  
 Each sample is conditioned at (21 ± 5) ° C and (85 ± 5)% relative humidity for at least 4 hours.  
 The test is performed with the inside of the mask in contact with the bacterial suspension.  
 The test area has a size of 49 cm<sup>2</sup>.  
 The flow rate during the test is equal to 28.3 l/min.

The final test value is the lowest BFE result found in the tests performed.

(\*\*) Sampling not subject to ACCREDIA accreditation

(<sup>1</sup>) Information provided by the customer, the laboratory declines all responsibility.

---

## FOLLOWS TEST REPORT N° 5\_24/03/21

---

Issue date 24/03/2021

### Legislative notes

(14683en) = EN 14683: 2019 Facial masks for medical use - Requirements and test methods - Table 1 "Performance requirements for masks for medical use".

I = Type I medical face mask  
II = Type II medical face mask  
IIR = Type IIR medical face mask

### Declaration of Conformity

For the parameters analyzed, according to the EN 14683: 2019 Table 1 standard, the sample complies with the performance characteristics envisaged for TYPE II medical masks.

The results contained in this Report refer exclusively to the sample as received in the laboratory

The results refer to the tested sample only and do not imply a lot or whole lot approval; if the Customer is responsible for the Sampling phase, the results refer to the sample received. The Laboratory declines all responsibility for the calculated results considering the sampling data provided by the Customer.

The samples are kept in this laboratory until the completion of the tests, excluding the official samples.

The uncertainties associated with the test results were calculated with a coverage factor  $k = 2$  equal to a confidence level of 95%. In the event that a declaration of conformity is formulated, for the purposes of the acceptability of the analytical data with respect to a limit value / guide value, the estimated uncertainty and / or estimated confidence interval is not taken into account.

It is absolutely forbidden to modify even partially the data contained.

U.M =Unit of measure LOQ =Quantification limit Ref.=Normative reference PP=Internal method (Test procedure)

Total or partial reproduction of this copy is prohibited, unless authorized in writing by the laboratory.

----- **End of the test report** -----

**Technical Director**  
Dott. Giuseppe Mazza

Document digitally signed by Dr.  
Giuseppe Mazza - Order of Chemists  
of Campania N.1147